

# DECLARATION OF CONFORMITY

SILVERTRONIC Ltd. confirm that the products listed below have been Designed, Manufactured and Tested to, and therefore are fully compliant to;  
**IEC1010-1 1990 inc. Amendment 2-1995 & IEC1010-2-031 First Edition-1993**  
and thus meet **THE LOW VOLTAGE DIRECTIVE (LVD) 73/23/EEC and amending directive 93/68/EEC**  
**GS38 Compliant**

PRODUCT TYPE / DESCRIPTION				GS38 FUSED TEST LEAD AND CROC. CLIP KIT	
PART No.				131009/RA	
IEC1010 RATING	VOLTAGE	* SEE BELOW	Volts	INSULATION CATEGORY	III ( 3 )

- \* Fused Probes (121204/RA) Rated at 1000V CAT III
- \* Flanged Croc. Clips (126011) Rated at 600V CAT III

MAXIMUM RATED CURRENT	*0.5 as fuse	<b>Amps</b>
POLLUTION DEGREE	2	
INSULATION CATEGORY	DOUBLE/REINFORCED	

INCLUDES FOLLOWING ASSEMBLIES / COMPONENTS SUBJECT TO INDIVIDUAL TEST.	DESCRIPTION	PART No.
	FUSED TEST LEAD ASSEMBLY	121204/RA
	FLANGED SCREW-ON CROC CLIP	126011

\* Current limited to fuse fitted



All Testing carried out by Third Party Testhouse  
**CRANAGE EMC TESTING Ltd.**  
( LVD TESTING DEPARTMENT )



www.silvertronic.com  
e-mail: sales@silvertronic.com



Registration No. 1993292

**North American Office**  
4405 Sharon Copley Road,  
Medina, Ohio, 44256.  
Toll Free: (866) 556 3308  
Facsimile: (330) 722 2281



This drawing is not to be disclosed to a third party without the authority from the company.©

Scale N.T.S.  
IEC 1010 COMPLIANT

121204/RA/R & B LEAD ASSEMBLIES



126011R & B SCREW-ON CROCS



1000V CAT III IEC 1010  
FUSED 500mA/1000V/50KA CE

ASSEMBLY JIG NUMBERS	-----J-----J-----
	-----J-----J-----

PACKAGING	119502
-	-

PRINTING AS ILLUSTRATED
BOTH PROBES ONE SIDE ONLY

TOLERANCES	CABLE LENGTHS +/- 5%
------------	----------------------

DO NOT SCALE.  
ALL DIMENSIONS IN mm.